



CONTENTS

No. of ordered detectors: **1**
Product: **Radtrak2**
Order ID:
Account:
Measurement period: **90 - 365 days**
Deploy By: **Date on Detector**

LOGIN CREDENTIALS

Commission number:
Password:

Radtrak²® | Long-Term Radon Test for Dwellings

INSTRUCTION FOR LONG-TERM RADON MEASUREMENT

- 1 Read these instructions in their entirety. The detector(s) should be placed accordingly, with the measurement period being at least 90 days, not exceeding 365 days. The measurement begins when the radon-proof plastic bag is opened. Immediately place device.
- 2 Go to www.radonline.com and use your commission number and password above to log into **MyPages**. Alternatively, scanning the QR code provided above with any smart device will take you to the commission and log in automatically. iOS users (11+) can scan the code with their camera. Android devices will need an app, QR Scanner or similar. Fill in measurement information, including specific detector location, detector number, and start date. The detector(s) should remain deployed in the same location until they are ready to be returned.
- 3 Upon completion of test, log back into **MyPages** and provide the measurement end date. Confirm your location information and Save. Package up your detector(s) and return immediately to the laboratory address below. Retain this document for your records.
- 4 Upon receipt your detectors will be analyzed. Measurement Reports are typically available within 5-10 business days. The report can be emailed to the email address provided, and may be retrievable from the website as well.

RECOMMENDED DETECTOR PLACEMENT

Please note! Some states or organizations may have differing recommendations. If you have a question regarding placement, contact your state radon agency.

- At least 20" above the floor
- At least 3' from windows or exterior doors
- At least 4" from objects or outside walls
- **Do not place on any mantels or other natural stone**
- The EPA recommends the test kit be placed in the lowest, lived-in level of the home (i.e. a frequently used basement)
- It should be put in a room that is regularly used such as a living room, den, or bedroom, **but not in a kitchen or bathroom**

The Radtrak²® radon detector is manufactured and analyzed by Radonova Laboratories AB, the global leader in radon measurements. This device is designed for long-term testing per EPA protocol, or 90 days to one year. Long-term detectors provide a more reliable representation of your average radon levels, and are less vulnerable to under/over-reporting due to fluctuations in radon levels.

RETURN DETECTOR(S) TO

Radonova, Inc.
900 Oakmont Lane Suite 207, Westmont IL 60559

QUESTIONS REGARDING THE MEASUREMENT

Email: help@radonova.com
Telephone: 331.814.2200